

# Certificate of Analysis



**YSI Incorporated**  
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA  
Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

Product Category                      Bioanalytical Standard  
Product Model Number                      2790  
Lot Number                                      19K10825  
Method    Comparative enzymatic testing of Production standard vs. Metrology standard

Chemistry Tested	Determination	Tolerance	Result
ETHANOL STD A 2.0G/L	(Average of production) - (average of metrology)	±40 mg/L	7 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	0.5%
ETHANOL STD B 3.2G/L	(Average of production) - (average of metrology)	±64 mg/L	9 mg/L
	Coefficient of variation for production standard (CV, %)	≤2%	1.2%

Conforms to acceptance criteria

Completed by: Lucas Stover    Date: 11/08/2019



Lisa M. Abel  
Director, Quality and Corporate  
Responsibility